



Certificate of Calibration

Company:
Address:

Date Certified: April 17, 2017
Due Date: April 17, 2018
Technician: Sample

Make: Watts / Orange Research **Model:** TK99E **Serial Number:** Sample
Range: 0-15 PSI **Service:** Backflow Test Kit **Tag Number:** Sample

Applied Pressure (PSI)	Indicator (As Found)	Indicator (As Left)
14.0 PSI	14.2 PSI	14.2 PSI
10.0 PSI	10.2 PSI	10.2 PSI
5.0 PSI	5.0 PSI	5.0 PSI
2.0 PSI	2.0 PSI	2.0 PSI
1.0 PSI	1.0 PSI	1.0 PSI
0.0 PSI	0.0 PSI	0.0 PSI

Comments: Calibrated and pressure tested to 160 PSI.

This unit has been calibrated using equipment and standards traceable to the National Research Council of Canada (NRC) or the National Institute of Standards and Technology (NIST).

Calibration Standard(s) Used

Asset Used	Model Number	Serial Number	Asset Due Date
Crystal	Sample	Sample	February 17, 2018

Sample
Technician

Date