



**CERTIFICATE OF CALIBRATION**

**ORDER NO** 48335 **AMBIENT TEMPERATURE** 20 C  
**SERIAL NUMBER** 4311 **HUMIDITY** 10%  
**DIAL SIZE** 2" **MANUFACTURER** FAIRBANKS **WEIGHT** SN# 60041  
**PRESSURE RANGE** 0-20,000 PSI **TEST MEDIUM** OIL  
**SUB DIVISIONS** 1000 PSI **MAXIMUM ERROR** - 60 PSI  
**CONNECTION** 1/4" NPT **DATE** JANUARY 9-2017  
**INSPECTOR** ERIC TURCOTTE **LOCATION TESTED:** 9535 42 AVE  
EDMONTON, ALBERTA, T6E 5R2

THIS GAUGE IS CERTIFIED TO MEET AND/OR EXCEED THE MANUFACTURERS SPECIFICATIONS, AS OUTLINED IN CUSTOMER STANDARD #038, DOCUMENT CS-038-1, ACCURACY +/- 1.0% FS.

<u>TESTED GAUGE</u>	<u>DEADWEIGHT</u>	<u>UPSCALE ERROR</u>
0	0	0
5000 PSI	5000	- 60 PSI
10,000 PSI	10,000	- 60 PSI
15,000 PSI	15,000	- 60 PSI
20,000 PSI	20,000	- 60 PSI

Technician:  Sheldon Cunningham Q.C. Manager  Bo Svoboda

INVOICE: PCL-039167//SJ

\*\*ALL TEST EQUIPMENT IS TRACEABLE TO N.I.S.T. STANDARDS

CUSTOMER STANDARD DOCUMENT #CS-038

CUSTOMER IN SERVICE DATE: \_\_\_\_\_

NEXT CALIBRATION DUE DATE: \_\_\_\_\_

SERVICE TECHNICIAN (SIGNATURE) \_\_\_\_\_

SERVICE TECHNICIAN: (PRINT) \_\_\_\_\_